

ADVISORY NOTICE

LAURA XL software upgrade – version 1.2.0.20

LAURA XL, Catalogue No.: INS00065

Dear customer,

We would like to inform you, that latest released software upgrade (version 1.2.0.20) for automatic urine analyser LAURA XL includes these improvements and changes:

- upgraded automatic recognition of urine sediment – improvement of automatic recognition algorithm ensures more accurate and more sensitive results of urine sediment analysis
- manual editing of chemistry results, which enables user to modify test result according to his expertise
- improvements in data security
- repair of some small bugs from previous versions

In case of any questions, please contact your local ERBA team.

Best regards,

Erba Lachema s.r.o.

Technical Product Management

Brno, 02nd Juny 2022



Erba Lachema s.r.o.

Karásek 2219/1d, 621 00 Brno

DIČ: CZ26918846 -7-