

DENSILAMETER® II



Comfortable Determination
of Microbial Suspension Density



DENSILAMETER® II

Patented Method of Inoculums Measurement

- Elimination of subjective errors related to determination of density of inoculums
- Economic efficiency
- Measurement accuracy
- Measurement of turbidity from 0.0 to 15.0 McFarland
- Two production calibrations - user calibration as optional
- Variability of user test tubes
- Low volume of suspension needed



Specifications are subject to change without prior notice.
Pictures used only for representation purposes- subject to change without prior notice.



Erba Lachema s.r.o.

Karásek 1d, 621 00 Brno, Czech Republic

Tel.: +420 517 077 111, fax: +420 517 077 077

E-mail: diagnostics@erbalachema.com

www.erbalachema.com



Contact info / map