

## ADVISORY NOTICE

38\_2023\_BILIRUBIN TOTAL BIL T 330

Dear customers,

With a mission to continuously enhance user experience and satisfaction, we would like to inform you regarding the change of storage temperature (which is valid also for transport conditions) of diagnostic reagent BILIRUBIN TOTAL (BIL T 330, Cat. No. XSYS0023) from 2–25 °C to 2–8 °C.

Changing of storage temperature for BILIRUBIN TOTAL was performed to unify storage conditions of all Erba products in diagnostic reagents group. If you accidentally store the BILIRUBIN TOTAL diagnostic reagent according to the original storage conditions, this will not affect its performances or shelf life.

In case of any questions, please do not hesitate to contact your local ERBA team.

Best regards,

Erba Lachema s.r.o.  
Technical Product Management

Brno, 9<sup>th</sup> November 2023



Erba Lachema s.r.o.  
Karásek 2219/1d, 621 00 Brno  
DIČ: CZ26918846 -7-

Ref. No. AN\_38\_2023