

ADVISORY NOTICE

14_2024_UREA 250

Dear customer,

We hereby inform you that reagent **UREA 250, LOT 2302131** (reg. n. 12000079, cat. n. BLT00061) which is used for quantitative *in vitro* determination of Urea in human serum, plasma and urine, was found to show unacceptable results of calibrations and quality control measurements.

If you observe any issue with this LOT during calibration or quality control measurement (such as blank/concentration/factor out of range), please stop your analysis immediately, do not continue with other measurements with the same LOT (especially patient samples) and claim this product in our CRM system (support.erbamannheim.com). Once we receive your claim, we will provide you a replacement free of charge.

In case you have any additional question, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.



Erba Lachema s.r.o.
Karásek 2219/1d, 621 00 Brno
DIČ: CZ26918846 -7-

Brno, 31st May 2024

Ref. No.AN_14_2024