

## ADVISORY NOTICE

34\_2024\_ Re-assignment calibrators and controls for CK-MB

Dear customer,

We hereby inform that the target values for all calibrator and control lots listed below for the CK-MB method have been reassigned.

The reassignment was carried out based on the verification of the calibrator's traceability with the IFCC reference method for the determination of CK-MB.

New target values in calibrators XL MULTICAL:

XL MULTICAL		CKMB	
Lot	Expiry	U/L	$\mu\text{kat/L}$
2310071	2025-04	482	8,03
2310072	2025-04	482	8,03
2404135	2026-01	460	7,67
2405057	2026-01	460	7,67
2405058	2026-01	460	7,67
2407121	2026-03	434	7,23
2409055	2026-03	434	7,23

New target values in controls ERBA:

Control material			CKMB					
Name	Lot	Expiry	U/L			$\mu\text{kat/L}$		
			value	interval	SD	value	interval	SD
ERBA NORM	2302119	2025-12	38,2	28,6-47,8	3,2	0,64	0,49-0,79	0,05
ERBA NORM	2302121	2025-12	38,2	28,6-47,8	3,2	0,64	0,49-0,79	0,05
ERBA NORM	2310172	2026-06	37,3	28,0-46,6	3,1	0,62	0,47-0,77	0,05
ERBA PATH	2302120	2025-12	85,8	64,2-107,4	7,2	1,43	1,07-1,79	0,12
ERBA PATH	2302122	2025-12	85,8	64,2-107,4	7,2	1,43	1,07-1,79	0,12
ERBA PATH	2310173	2026-07	77,5	58,0-97,0	6,5	1,29	0,96-1,62	0,11

In case you have any additional question, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.  
Technical product management

**Erba Lachema s.r.o.**  
Karásek 2219/1d, 621 00 Brno  
DIČ: CZ26918846 -7-

Brno, 4<sup>th</sup> November 2024

Ref.No.AN\_34\_2024

